

**Tablet Dis-integration Test Apparatus
(Common Motor & Timer for Two Basket)
Model: μ TDTCa110 | Make: ANALAB | HSN Code: 90318000**



Supplied with 1000 ml Glass Beaker – 02 Nos., 6 tube basket assembly – 02 Nos., Guided Discs – 12 Nos., Temperature Probe – 02 Nos., Power Cord, Operational & Instruction Manual with SOP, Calibration Test Certificate (Without Traceability Document) and Warranty Card.

Technical Specification:

Stroke Length	: 55mm ± 2mm
Number of Stroke Frequency Rate	: 29–32 Cycles Per Min.
Readable Temperature Range	: 10.0°C to 50.0°C
Temperature Control Range	: 30.0°C to 40.0°C
Temperature Accuracy	: ± 0.2°C
Temperature Resolution	: 0.1°C
Programmable Timer	: 1 sec. to 9 hrs.: 59 min: 59 sec.
Heating time to 37°C	: Less than 20 min.
Heating Technology	: Bathless
Heater Capacity	: DC Heater-150 Watt
Heating	: Bathless heating results into fast heating and save water
Temperature Sensor	: PT-100
Number of Beaker	: 02 Nos. – Capacity 1000 ml
Number of Motor	: 01 Nos.
No of Timer	: 01
Number of Basket	: 02 Nos. (06 Test Tube Per Basket)
Control of Basket	: Common Motor & Timer for Two Basket
Display	: 40 x 4 Line Alphanumeric Backlit LCD
Loading of Basket	: Magnetised for ease of cleaning.
Illumination	: White Light for better visibility of dis-integration process
Parking of Basket	: Automatic after completion of test
Keypad	: Membrane waterproof polycarbonate soft touch keypad compliance to UL/ROHS
Pharmacopeia Complies	: As per USP, IP, Ph. Eur.
Date	: 12/24 Hours Format
Time	: Real time clock
Power Supply	: 230VAC ± 10%, 50 Hz
Dimensions	: 460 x 450 x 710 mm (L x W x H)
Weight	: 19.5 Kgs (Net) & 24.0 Kgs (Gross)

Optional

1. IQ/OQ/PQ/DQ Documents

Note: a) Please provide Customer's Name & Address prior to dispatch.

b) To perform onsite documentation, visit charge will be extra depending upon place of Installation.

c) Traceability documents along with installation report.

d) Standards / Chemicals / Buffers required to perform documents are in customer's scope.

e) We don't send documents by email.

2. Keypad Screen Guard

3. Number of Basket (Type B): 02 Nos. (03 Test Tube Per Basket)

(For Diameter of Tablet More than 18mm) available Optionally.

4. GMP Enclosure made up of Stainless Steel

Product Range

•pH/mV/°C/ORP Analyzer •Conductivity/TDS/Resistivity/Salinity/°C Analyzer
•Multiparameter Analyzer (pH/ORP/Conductivity/TDS/Resistivity/Salinity/°C)
•Karl Fischer Titrator •Auto Titrator •Automatic Melting Point Apparatus
•Flame Photometer •Friability Test Apparatus •Tablet Dis-integration Test Apparatus
•Tap Density Test Apparatus •Tablet Hardness Tester •Leak Test Apparatus
•Tablet Dissolution Test Apparatus •Electromagnetic Sieve Shaker •Ultrasonic Bath (Sonicator)
•Analytical Electronics Balance •Halogen Based Moisture Analyzer
•Visible & UV-Vis. Spectrophotometer •Nitrogen Gas Generator •Zero Air Gas Generator
•Combination Nitrogen / Zero Air Gas Generator •Air Drying Unit for TOC Analyzer
•Dew Point Apparatus •Gas Handling & Purification System •Laboratory Furniture & Fume Hood
•Exhaust Hood for AAS/ICP.

ANALAB SCIENTIFIC INSTRUMENTS PVT. LTD.

AN ISO 9001 : 2015 CERTIFIED CO.
CIN: U33110GJ2005PTC046555



Plot No. 361, Sayajimarket,

N. H. No. 8 By-Pass,

At & Post : Sayajipura,

Vadodara - 390019, Gujarat, India.



+91 265 2514404, 9714503393



info@analab.co.in



www.analab.co.in, www.analabscientific.com

