

Microcontroller Based Flame Photometer

Model: μ FlameCal5 | Make: ANALAB | HSN Code: 90275010



Supplied with Two Filters (Na and K), Air Supply Unit, Atomizer Tube, Sample Beaker (06 Nos.), Gas Lighter, LPG Tube, Power Cord, Operational & Instruction Manual with SOP, Calibration Test Certificate (Without Traceability Document) and Warranty Card.



Technical Specification:

Element Range	: Na, K, & Li: 1-100 ppm, Ca: 15-100 ppm
Sensitivity	: Na: 0.5 ppm, K: 0.5 ppm, Li: 0.5ppm, Ca: 15 ppm
Filter Selection	: Automatic by soft touch key
Detector	: Silicon Photodiode
Linearity	: Better than 2%
Unit of Measurement	: ppm
Reproducibility	: <1% Coefficient of Variation (CV) for 20 consecutive samples using 10 ppm Na set as maximum standard
Quick Calibration Facility	: With highest standard
Calibration	: One standard per Element
Calibration Data Storage	: Only One element calibration data can be stored.
Ignition System	: Manual
Flame Type	: LPG & Dry Oil Free Air
Gas Control	: Adjustable with regulator
Atomizer	: Axial Flow Type
Error Messages	: “Home Detection” .
Display	: 16x2 Line Alphanumeric LCD with Backlit
Keypad	: Membrane waterproof polycarbonate soft touch keypad compliance to UL/ROHS
Air Supply Unit (Separate Unit)	: Consist of Air Compressor, Pressure Gauge, Pressure Regulator, Moisture Filter, PU Tube (2 Meters) and Air filter with moisture absorbent catalyst.
Calibration Data Storage	: Non-Volatile Memory, which can store, last calibration data for only one element.
Power Supply	: 230VAC \pm 10%, 50 Hz

Optional

1. IQ / OQ / PQ / DQ Documents

Note: a) Please provide Customer's Name & Address prior to dispatch.

b) To perform onsite documentation, visit charge will be extra depending upon place of Installation.

c) Traceability documents along with installation report.

d) Standards / Chemicals / Buffers required to perform documents are in customer's scope.

e) We don't send documents by email.

2. Calcium (Ca) Filter – HSN Code : 90022000

3. Lithium (Li) Filter – HSN Code : 90022000

4. Sodium 100 ppm Stock Solution – 500 ml - HSN : 38220090

5. Potassium 100 ppm Stock Solution – 500 ml - HSN : 38220090

6. Calcium 100 ppm Stock Solution – 500 ml - HSN : 38220090

7. Lithium 100 ppm Stock Solution – 500 ml - HSN : 38220090

8. Keypad Screen Guard – HSN Code : 39199010

9. Traceability Documents

Available other Models of Flame Photometer

- Microcontroller Based Flame Photometer Model: μ FlameCal5 with Single Standard Calibration
- Microcontroller Based Flame Photometer Model: μ FlameCal10+ with Three Standard Calibration as per Non-GLP Format
- Microcontroller Based Flame Photometer Model: μ FlameCal50 with Five Standard Calibration with Printing Facility as per GLP Format
- Microcontroller Based Flame Photometer Model: μ FlameCal100 with 21 CFR Part-11 Compliance & Printing Facility as per GLP Format

Product Range

- pH/mV/°C/ORP Analyzer • Conductivity/TDS/Resistivity/Salinity/°C Analyzer
- Multiparameter Analyzer (pH/ORP/Conductivity/TDS/Resistivity/Salinity/°C)
- Karl Fischer Titrator • Auto Titrator • Automatic Melting Point Apparatus
- Flame Photometer • Friability Test Apparatus • Tablet Dis-integration Test Apparatus
- Tap Density Test Apparatus • Tablet Hardness Tester • Leak Test Apparatus
- Tablet Dissolution Test Apparatus • Electromagnetic Sieve Shaker • Ultrasonic Bath (Sonicator) • Analytical Electronics Balance • Halogen Based Moisture Analyzer
- Visible & UV-Vis. Spectrophotometer • Nitrogen Gas Generator • Zero Air Gas Generator • Combination Nitrogen / Zero Air Gas Generator • Air Drying Unit for TOC Analyzer • Dew Point Apparatus • Gas Handling & Purification System • Laboratory Furniture & Fume Hood • Exhaust Hood for AAS/ICP.

ANALAB SCIENTIFIC INSTRUMENTS PVT. LTD.

AN ISO 9001 : 2015 CERTIFIED CO.
CIN: U33110GJ2005PTC046555

Plot No. 361, Sayajimarket,
N. H. No. 8 By-Pass,
At & Post : Sayajipura,
Vadodara - 390019, Gujarat, India.

+91 265 2514404, 9714503393

info@analab.co.in

www.analab.co.in, www.analabscientific.com

