

Multiparameter Analyzer (pH/ORP/Conductivity/TDS/°C)

Model: μ MultiCal10 | Make: ANALAB | HSN Code: 90278990



Supplied with Combination Glass pH Electrode (0-70°C), Conductivity Cell K=1.0, Standard Buffer Tablets, Temperature Probe, Electrode Stand consists of Clamp, Dust Cover, DC Adaptor, Operational & Instruction Manual with SOP, Calibration Test Certificate (Without Traceability Document) and Warranty Card.

Technical Specification:

pH	
Range	: -2.000 to 20.000 pH
Selectable Resolution	: 0.1 / 0.01 / 0.001 pH
Relative Accuracy	: ± 0.002 pH
mV/ORP	
Range – mV/RmV	: ± 2000.0
ORP	: EH auto calibration to standard hydrogen electrode
Resolution	: 0.1
Relative Accuracy	: ± 0.2 mV or 0.05%, whichever is greater.
Slope	: 80 - 120 %
Auto Buffer Recognition	: 1) 6.86 / 7.00 & 4.01 / 9.18 / 10.01 pH @ 25°C for Two points 2) 4.01, 6.86 / 7.00 & 9.18 / 10.01 pH @ 25°C for Three points
Calibration Reminder Alarm	: 0-99 hours
Result	: Auto / Manual
pH Buffer Standard	: US / NIST / EURO / DIN
Short Link	: Provided to diagnose the problem either with Instrument OR Electrode.
Conductivity Range	: 0 - 20.00, 20 - 200.0, 200 - 2000 μ S/cm, 2 - 20.000, 20 - 200, 200 - 2000 mS/cm for K=1.0 : 0 - 10.00, 10 - 100.0, 100 - 1000 μ S/cm, 1 - 10.000, 10 - 100, 100 - 1000 mS/cm for K=0.5 : 0 - 2.00, 2 - 20.0, 20 - 200, 200 - 2000 μ S/cm, 2 - 20, 20 - 200 mS/cm for K=0.1 (Range value change depending upon cell selection)
Conductivity Resolution	: 0.001, 0.01, 0.1 μ S/cm & 0.001, 0.01, 0.1, 1 mS/cm
TDS Range	: 0 to 10, 100, 1000, 10000 ppm, 100ppt, 1000ppt (Depending on TDS Factor)
TDS Resolution	: 0.001, 0.01, 0.1 ppm & 0.001, 0.01, 0.1 ppt (1ppt= 1000ppm)
Number of Range	: Six
Calibration Standard	: Six Conductivity & TDS Standard - One Per Range
Calibration Factor	: Six - One Per Range
Range Selection	: Auto / Manual
Suitable Cell Constant	: 0.1, 0.5 & 1.0 K
Adjustable Cell Constant	: 0.09 to 0.11 K, 0.45 to 0.55 K & 0.90 to 1.1 K

Conductivity to TDS Conversion :	0.40000 to 1.00000 (Adjustable Factor)
Temperature Coefficient :	Six - 0-10.000 % per °C (Adjustable) - One Per Range
Reading :	Conductivity at actual temperature and at 25°C & TDS with scrolling display.
Temperature Range :	-5.0 to 130.0°C
Temperature Resolution :	0.1°C
Temperature Compensation :	Auto/Manual
Temperature Calibration	: In-built software and can be calibrated by users
Calibration & Analysis	: pH Mode, ORP Mode, Conductivity Mode and TDS mode.
Error Message	: “Wrong standard / Cell”, “Cell Short”, “Cell Open”, “Try Upper Range”, “Try Lower Range”, “Bad Buffer / Electrode”, “pH Over Range”, “Temperature Probe Error”, “In-complete Calibration”, “Calibration Due”, & “Low Memory”
Display :	16 x 2 Line Alphanumeric Backlit LCD.
Keypad :	Membrane waterproof polyester soft keypad compliance to UL/ROHS.
Calibration Data Storage :	Non-Volatile Memory can store, print & view calibration details.
Analysis Data Storage :	More than 500 analysis data which can be viewed & Print, company name, analysis date & time, record number, *Username, *Sample Name, *Bath Number/ID, , pH, mV, temperature reading, calibration data with date & time, conductivity at 25°C and conductivity at actual temperature, % Tempco, conductivity to TDS conversion factor, TDS, instrument serial number, Instrument Make & Model & *Signature etc. Secondly , analysis data can be print online without saving analysis data in Instrument. (*: Not Possible to enter Data by Instrument Keypad but it will be Handwritten)
Searching of data :	With record number by scrolling.
Report Formation	: As per Non-GLP Compliance.
Printer Interface :	Centronics Parallel Port/ Serial (RS-232 C) Port
Printer :	40 / 80 Column Dot Matrix Printer (Printer to be procure optionally)
Power Supply :	DC Adaptor

Optional

1. IQ/OQ/PQ/DQ Documents

Note:

- Please provide Customer's Name & Address prior to dispatch.
- To perform onsite documentation, visit charge will be extra depending upon place of Installation.
- Traceability documents along with installation report.
- Standards / Chemicals / Buffers required to perform documents are in customer's scope.
- We don't send documents by email.

2. 40 / 80 Column dot matrix printer

3. Keypad Screen Guard

Other Models of Multiparameter Analyzer

- Multiparameter Analyzer (pH/ORP/Cond./TDS/°C) Model: μ MultiCal10 with Printing Facility
- Multiparameter Analyzer (pH/ORP/Cond./TDS/Resistivity/Salinity/°C) Model: μ MultiCal50 with Printing Facility as per GLP Format
- Multiparameter Analyzer (pH/ORP/Cond./TDS/Resistivity/Salinity/°C) Model: μ MultiCal100 with 21 CFR Part-11 Compliance
- Handheld Multiparameter Analyzer (pH/ORP/Conductivity/TDS/°C) Model: μ MultiCalHandheld

Product Range

- pH/mV/°C/ORP Analyzer
- Conductivity/TDS/Resistivity/Salinity/°C Analyzer
- Multiparameter Analyzer (pH/ORP/Conductivity/TDS/Resistivity/Salinity/°C)
- Karl Fischer Titrator
- Auto Titrator
- Automatic Melting Point Apparatus
- Flame Photometer
- Friability Test Apparatus
- Tablet Dis-integration Test Apparatus
- Tap Density Test Apparatus
- Tablet Hardness Tester
- Leak Test Apparatus
- Tablet Dissolution Test Apparatus
- Electromagnetic Sieve Shaker
- Ultrasonic Bath (Sonicator)
- Analytical Electronics Balance
- Halogen Based Moisture Analyzer
- Visible & UV-Vis. Spectrophotometer
- Nitrogen Gas Generator
- Zero Air Gas Generator
- Combination Nitrogen / Zero Air Gas Generator
- Air Drying Unit for TOC Analyzer
- Dew Point Apparatus
- Gas Handling & Purification System
- Laboratory Furniture & Fume Hood
- Exhaust Hood for AAS/ICP.

ANALAB SCIENTIFIC INSTRUMENTS PVT. LTD.

AN ISO 9001 : 2015 CERTIFIED CO.
CIN: U33110GJ2005PTC046555

-  Plot No. 361, Sayajimarket,
N. H. No. 8 By-Pass,
At & Post : Sayajipura,
Vadodara - 390019, Gujarat, India.
-  +91 265 2514404, 9714503393
-  info@analab.co.in
-  www.analab.co.in, www.analabscientific.com

