

**Digital pH/mV Meter****Make: ANALAB Model: pHCal****Installation prerequisites**

➤ **Environment Conditions :**

The Instrument can operate within temperature and relative humidity conditions of 0 °C to 55°C and 5 to 95 % respectively. However an environment comfortable for human habitation (reasonable constant temperature and humidity conditions) is recommended for optimum performance and instrument lifetime.

Exposure to corrosive substances (weather gas, liquid or solid) may affect materials and components used in instrument and should avoid.

➤ **Power Supply Requirements:**

- 1) Stabilized Power supply 230 VAC  $\pm$  10 %, 50Hz.  
(Earthing voltage should be **Less than 3V AC** – between Earthing & Neutral)
- 2) One 5 Amp. Points require for Main unit

➤ **Chemicals and Accessories Required:**

- 1) Double Distilled Water for buffer preparation - 1 liter.
- 2) Wash Bottle for distilled water
- 3) Bottles to store buffer solutions. – Capacity 100 ml - (2 Nos.)
- 4) Volumetric flask / Beaker – Capacity -100 ml - (3 Nos.)
- 5) Tissue Paper Roll

**Note :**

For further clarification / support please email us at [service@analab.co.in](mailto:service@analab.co.in) OR  
Contact us on +91-9714503393 / +91-9824030428  
Contact Person: Mr. Rajesh Patel